

EC REP

LifeGlobal Europe, Rue de la Presse 4, 1000 Brussels Belgium T: 32-2 227 1129 F: 32-2 218 3141

1: 32-2 227 1129 F: 32-2 218 314 LifeClobal Group, LLC, 303 Sound



LifeGlobal Group, LLC, 393 Soundview Rd, Guilford, CT 06437 US T: 1-800-720-6375 F: 1-519-826-6947 Intl.: 001-519-826-5800 sales@LifeGlobal.com www.LifeGlobalGroup.com

(()

Certification of Analysis for global® total® w/ HSA

(Catalogue Numbers: HGGT-030, HGGT-060, HGGT-100)

Lot Number: HGGT-171120C

Expiry Date: 2018-01-29 yyyy-mm-dd

Production Date: 2017-11-20

yyyy-mm-dd

Indications for Use: Culture of human embryos from zygote to blastocyst, embryo transfer.

Caution: Federal Law (USA) restricts this device to sale by or on the order of a physician (or properly licensed practitioner).

This product contains human serum albumin, a derivative of human blood. The human serum albumin used in the preparation of this product has been heated at 60°C for ten hours.

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and HTLV I/II and non-reactive for HbsAg, HCV RNA and HIV-1 RNA and syphilis. No known test methods can offer assurances that products derived from human blood will not transmit infectious agents.

global® total® w/ HSA contains the antibiotic gentamicin sulfate. Appropriate precautions should be taken to ensure that the patient is not sensitized to this antibiotic.

Discard unused medium within 7 days of opening. Do not use after expiry date. Devices are not intended for resterilization.

Instructions for Use: See global® total® w/ HSA Instructions for Use.

Storage Requirements: Store at 2-8°C and protected from light. When stored unopened at 2-8°C and protected from light, the product is stable until the expiration date shown on the bottle label.

Raw Materials

Chemicals used by LifeGlobal meet ASC, USP, or NF standards.

We certify that this product has been manufactured according to current Good Manufacturing Practice (cGMP). The quality of this particular batch was tested and evaluated by stringent quality control procedures. On the basis of superior test results, the batch was considered suitable for the use intended.

Assay (performed for each batch)	Specification	Results	
Physicochemical Tests			
pH (with 5% CO ₂)	7.2-7.4	7.27	
Osmolality	260-270 mOsM	265	mOsM
Biological Tests			
Endotoxin (LAL)	≤ 0.5 EU/ml	< 0.03	EU/ml
Sterility Test (bacterial and fungal screen, SAL 10 ⁻³)	PASS	PASS	
Biological Assays			
1-cell Mouse Embryo Assay (% expanded blastocysts at 96 h of culture)	<u>></u> 80%	100	%