

# QUALITY ASSURANCE CERTIFICATE



Sperm VitalStain™

Optimised one-step staining technique for assessment of sperm vitality, one of the basic tools used in semen analysis. The technique is based on the principle that dead cells will take up the red colour of eosin stain. Nigrosin is used as a counter stain to facilitate visualization of the otherwise unstained sperm.

Containing: 2 x 10 mL

Catalogue no: SVS-010

Batch:  
**010SVSVH19**

Production date:  
**19 August, 2015**

Expiry date:  
**18 August, 2017**

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

**Quality Control procedures:**

**Accepted values:**

**Test results:**

***Physical analyses:***

*pH at room temperature in air*

7.3-9.5

Comply

***Biological analyses:***

*Batch to batch functional variation*

<5 %

Comply

Date of release: 25 August, 2015

*This document is electronically generated and does not bear a signature.*

**Paul V. Holmes**, MSc, PhD, DrMedSc  
C.E.O.

**Marina Danilova**,  
Quality Assurance Coordinator

Nidacon International AB  
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden  
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15  
E-mail: [contact@nidacon.com](mailto:contact@nidacon.com) Web: [www.nidacon.com](http://www.nidacon.com)