

QUALITY ASSURANCE CERTIFICATE



PureSperm[®] 100

Colloidal stock suspension optimised for discontinuous density gradient separation of human sperm.

Containing: 250 mL

Catalogue no: PS100-250

Batch:
250PS100XH10

Production date:
10 August, 2017

Expiry date:
09 August, 2019

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

Quality Control procedures:

Accepted values:

Test results:

Physical analyses:

pH at room temperature in air

7.4-7.8

Comply

Osmolality (mOsm)

300-310

Comply

Sterility and toxin analyses:

Microbiological growth control

No growth

Comply

*Endotoxin transfer during treatment,
quantitative LAL assay (EU/mL)*

< 1.00

Comply

Biological analyses:

Human sperm survival after 18 hrs

>70 %

Comply

Date of release: 18 September, 2017

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, DrMedSc
C.E.O.

Marina Danilova
Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com