



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

**LOT NUMBER:** 702IN203/IN303/IN353  
**PRODUCT NAME:** V-ICSIPET INJECTION 20°, 30°, 35°  
**REFERENCE:** V-INJ-S3-20,-30,-35  
**EXPIERY DATE:** 20.02.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 29.03.2017