

CERTIFICATE OF ANALYSIS

LOT NUMBER: 703HO20/HO30
PRODUCT NAME: V-iCSIPET HOLDING 20°, 30°
REFERENCE: V-HOLD-20,-30
EXPIERY DATE: 27.03.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 03.05.2017

