

CERTIFICATE OF ANALYSIS

LOT NUMBER: 704IN203/IN303/IN353
PRODUCT NAME: V-ICSIPET INJECTION 20°, 30°, 35°
REFERENCE: V-INJ-S3-20, 30, 35
EXPIERY DATE: 24.04.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 06.06.2017

