

## CERTIFICATE OF ANALYSIS

---

**LOT NUMBER:** 706IN203/IN303  
**PRODUCT NAME:** V-ICSIPET INJECTION 20°, 30°  
**REFERENCE:** V-INJ-S3-20, 30  
**EXPIERY DATE:** 26.06.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 14.08.2017

