



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 709HO20/HO30
PRODUCT NAME: V-ICSIPET HOLDING 20°, 30°
REFERENCE: V-HOLD-20,-30
EXPIERY DATE: 15.09.2020

| TEST | SPECIFICATION | RESULT |
|---|---------------|--------|
| Gamma Sterilization | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blactocyst in 96 h | ≥ 80 % | PASSED |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.10.2017