

QUALITY ASSURANCE CERTIFICATE



Sperm VitalStain™

An optimised staining technique for assessment of sperm vitality

Containing: 2 x 10 mL

Catalogue no: SVS-010

Batch:
010SVSZC22

Production date:
22 March, 2019

Expiry date:
21 March, 2021

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

Quality Control procedures:

Accepted values:

Test results:

Physical analyses:

pH at room temperature in air

7.3-9.5

Comply

Biological analyses:

Batch to batch functional variation

<5 %

Comply

Date of release: 17 April, 2019

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, DrMedSc
C.E.O.

Marina Danilova,
Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com