

QUALITY ASSURANCE CERTIFICATE



PureSperm® 40/80

For separation of human sperm with density centrifugation

Containing: 2 x 20 mL

Catalogue no: PSK-020

Batch:

020PS40YF25

020PS80YF26

Production date:

25 June, 2018

26 June, 2018

Expiry date:

24 June, 2020

25 June, 2020

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

Quality Control procedures:

Accepted values

Test results

PureSperm® 40

PureSperm® 80

Physical analyses:

pH at room temperature and in air

7.4–7.8

Comply

Comply

Osmolality (mOsm)

300-310

Comply

Comply

Sterility and toxin analyses:

Microbiological growth control

No growth

Comply

Comply

Endotoxins, transfer during

treatment, quantitative LAL assay (EU/mL)

< 1.00

Comply

Comply

Biological analyses:

Sperm survival 18 hrs after

density-gradient preparation

>70%

Comply

Comply

Date of release: 12 September, 2018

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, DrMedSc
C.E.O.

Marina Danilova
Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com