

QUALITY ASSURANCE CERTIFICATE

SpermCryoProtec™

Optimised for freezing human sperm

Containing: 2 x 20 mL

Catalogue no: SCP-020

Batch:
020SCPZF03

Production date:
03 June, 2019

Expiry date:
02 June, 2020

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

Quality Control procedures:

Accepted values:

Test results:

Physical analyses:

pH at room temperature in air

7.5-8.5

Comply

Sterility and toxin analyses:

Microbiological growth control

No growth

Comply

*Endotoxins, quantitative LAL assay
(EU/mL)*

< 1.00

Comply

Biological analyses:

*Recovery rate of the original motile
spermatozoa after freezing and thawing*

>50 %

Comply

Date of release: 14 June 2019

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, DrMedSc
C.E.O.

Marina Danilova,
Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com