

# QUALITY ASSURANCE CERTIFICATE



NidOil™

An oil overlay for use during ICSI or gamete and embryo culture, inside or outside the incubator.

Containing: 4x100 mL

Catalogue no: NO-400K

Batch:  
**100NOYA16**

Production date:  
**16 January, 2018**

Expiry date:  
**15 January, 2020**

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

## Quality Control procedures:

### **Physical analyses**

Density (d 20°C / 20°C, g/mL)

### Accepted values:

0.82-0.88

### Test results:

0.852

### **Sterility and toxin analyses**

Microbiological growth control

No growth

Comply

Endotoxins, quantitative LAL assay (EU/mL)

< 0.100

<0.005

### **Biological analyses**

Mouse Embryo Assay (% development to blastocyst stage at day 5)

≥80

Comply

Human sperm survival after 18 hrs

≥85

Comply

### **Peroxide analyses**

Peroxide level (mEq/kg)

<0.02

0.005

Date of release: 23 February, 2018

This document is electronically generated and does not bear a signature.

**Paul V. Holmes**, MSc, PhD, DrMedSc  
C.E.O.

**Marina Danilova**  
Quality Assurance Coordinator

Nidacon International AB  
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden  
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15  
E-mail: [contact@nidacon.com](mailto:contact@nidacon.com) Web: [www.nidacon.com](http://www.nidacon.com)