QUALITY ASSURANCE CERTIFICATE





An oil overlay for use during ICSI or gamete and embryo culture, inside or outside the incubator.

Containing: 4x100 mL Catalogue no: NO-400K

Batch: Production date: Expiry date:

100NOZB19 19 February, 2019 18 February, 2021

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

Quality Control procedures:		
	Accepted values:	Test results:
Physical analyses		
Density (d 20°C / 20°C, g/mL)	0.82-0.88	0.848
Sterility and toxin analyses		
Microbiological growth control	No growth	Comply
Endotoxins, quantitative LAL assay (EU/mL)	< 0.100	<0.005
Biological analyses		
Mouse Embryo Assay (% development to	<u>≥</u> 80	Comply
blastocyst stage at day 5)		
Human sperm survival after 18 hrs	<u>></u> 85	Comply
Peroxide analyses		
Peroxide level (mEq/kg)	<0.02	<0.0028

Date of release: 03 May 2019

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, DrMedSc Marina Danilova

C.E.O. Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com

Document ID TFC- 0868- 1

Published 2019-02-20