

QUALITY ASSURANCE CERTIFICATE



PureSperm® Buffer

Balanced salt solution optimised for the dilution of PureSperm®100 to prepare layers for density gradients.

Containing: 100 mL

Catalogue no: PSB-100

Batch:
100PSBZD11

Production date:
11 April, 2019

Expiry date:
10 April, 2021

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

Quality Control procedures:

Accepted values:

Test results:

Physical analyses:

pH at room temperature in air

7.4-7.8

Comply

Osmolality (mOsm)

300-310

Comply

Sterility and toxin analyses:

Microbiological growth control

No growth

Comply

*Endotoxins, quantitative LAL assay
(EU/mL)*

< 1.00

Comply

Biological analyses:

Human sperm survival after 18 hrs

>70 %

Comply

Date of release: 14 May 2019

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, Dr MedSc
C.E.O.

Marina Danilova,
Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel: +46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com