

QUALITY ASSURANCE CERTIFICATE

PureSperm[®] Wash

Optimised for washing the sperm pellet recovered from a PureSperm[®] density gradient separation, for use in the swim-up method and extending semen or sperm pellet for use in IUI.

Containing: 100 mL

Catalogue no: PSW-100

Batch:
100PSWVJ27

Production date:
27 October, 2015

Expiry date:
26 October, 2016

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

<u>Quality Control procedures:</u>	<u>Accepted values:</u>	<u>Test results:</u>
Physical analyses:		
pH at room temperature in air	7.3-8.5	Comply
Osmolality (mOsm)	290-300	Comply
Sterility and toxin analyses:		
Microbiological growth control	No growth	Comply
Endotoxins, quantitative LAL assay (EU/mL)	< 1.00	Comply
Biological analyses:		
Human sperm survival after 18 hrs	>70 %	Comply

Date of release: 07 December, 2015

This document is electronically generated and does not bear a signature.

Paul V. Holmes, MSc, PhD, DrMedSc
C.E.O.

Marina Danilova,
Quality Assurance Coordinator

Nidacon International AB
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15
E-mail: contact@nidacon.com Web: www.nidacon.com