

# QUALITY ASSURANCE CERTIFICATE



## PureSperm® Wash

for washing the sperm pellet recovered from a PureSperm® density gradient, for use in the swim-up method and for extending semen or sperm pellet for use in IUI.

Containing: 100 mL

Catalogue no: PSW-100

**Batch:**  
**100PSWZB25**

**Production date:**  
**25 February, 2019**

**Expiry date:**  
**24 February, 2020**

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

**Quality Control procedures:**

**Accepted values:**

**Test results:**

**Physical analyses:**

*pH at room temperature in air*

7.3-8.5

Comply

*Osmolality (mOsm)*

290-300

Comply

**Sterility and toxin analyses:**

*Microbiological growth control*

No growth

Comply

*Endotoxins, quantitative LAL assay  
(EU/mL)*

< 1.00

Comply

**Biological analyses:**

*Human sperm survival after 18 hrs*

>70 %

Comply

*Date of release: 06 May 2019*

*This document is electronically generated and does not bear a signature.*

**Paul V. Holmes**, MSc, PhD, DrMedSc  
C.E.O.

**Marina Danilova**,  
Quality Assurance Coordinator

Nidacon International AB  
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden  
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15  
E-mail: [contact@nidacon.com](mailto:contact@nidacon.com) Web: [www.nidacon.com](http://www.nidacon.com)