

# QUALITY ASSURANCE CERTIFICATE

SpermCatch™

Salt solution including hyaluronate, optimised for modulating sperm motility prior to ICSI.

Containing: 6 x 100 µL

Catalogue no: SC-100

Batch:  
**100SCWK22**

Production date:  
**22 November, 2016**

Expiry date:  
**21 November, 2017**

We certify that this product was manufactured according to set directives for medical devices. The quality of this particular batch was tested and evaluated objectively by stringent, quality control procedures.

<b>Quality Control procedures:</b>	<b>Accepted values:</b>	<b>Test results:</b>
<b>Physical analyses:</b>		
<i>pH at room temperature in air</i>	7.0-8.5	Comply
<i>Osmolality (mOsm)</i>	290-320	Comply
<b>Sterility and toxin analyses:</b>		
<i>Microbiological growth control</i>	No growth	Comply
<i>Endotoxins, quantitative LAL assay (EU/mL)</i>	< 1.00	Comply
<b>Biological analyses:</b>		
<i>Human sperm survival after 18 hrs</i>	>50 %	Comply

Date of release: 21 December, 2016

This document is electronically generated and does not bear a signature.

**Paul V. Holmes**, MSc, PhD, DrMedSc  
C.E.O.

**Marina Danilova**,  
Quality Assurance Coordinator

Nidacon International AB  
Flöjelbergsgatan 16B, SE-431 37 Mölndal, Sweden  
Tel:+46-31-703 06 30 Fax:+46-31-40 54 15  
E-mail: [contact@nidacon.com](mailto:contact@nidacon.com) Web: [www.nidacon.com](http://www.nidacon.com)