



*We care about your needs*

Halotech DNA SL  
C/ Faraday 7 , 28049 Madrid  
Spain  
Tel: +34 91 279 6950  
[info@halotech.es](mailto:info@halotech.es)  
[www.halotechdna.com](http://www.halotechdna.com)

### CERTIFICATE OF ANALYSIS

Product name: Halosperm®

Product number: HT-HS10

Lot Number: 1612


COMPONENTS	Parameter/Test	Specification	Result
Lysis Solution	pH	7±1	7
	Color	Colorless	Colorless
	Translucent appearance	Transparent, can be turbid	Transparent
	Aliquoted volume	110 ml	110 ml
Denaturant Agent	pH	1±1	1
	Smell	HCl	HCl
	Translucent appearance	Transparent, can be oxidiced	Transparent
	Aliquoted volume	1 ml	1 ml
Supercoated Slides	Translucent appearance	Transparent	Transparent
	Homogeneity	surface completely covered	surface completely covered
Eppendorf tube with Agarose	Color	Colorless	Colorless
	State	Solid	Solid
	Aliquoted volume	100 µl	100 µl

Results with biological sample, comparison two tests:

Accepted/Rejected (AC/RE)

AC

Expiration information	The expiry date is guaranteed to be valid for twelve months from the production date provided.
Production Date	10/06/2016
Expiration Date	10/12/2017
Certificate Issue date	20/06/2016



Francisco Arenal  
Quality Assurance Manager

**CSQ**  
**MED**  
ISO 13485



NOTICE: HalotechDNA products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from HalotechDNA.