

CERTIFICATE OF ANALYSIS

LOT NUMBER: 701IN302/IN303/IN352/IN353/IN203
PRODUCT NAME: V-ICSIPET INJECTION 20°, 30°, 35°
REFERENCE: V-INJ-S2-30/35, INJ-S3-20/30/35
EXPIRY DATE: 21.01.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 06.03.2017

