

CERTIFICATE OF ANALYSIS

LOT NUMBER: 705HO20/HO30
PRODUCT NAME: V-ICSIPET HOLDING 20°, 30°
REFERENCE: V-HOLD-20,-30
EXPIRY DATE: 29.05.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 06.07.2017

