

CERTIFICATE OF ANALYSIS

LOT NUMBER: 709IN203/IN302/IN303/IN352/IN353
PRODUCT NAME: V-ICSIPET INJECTION 20°, 30°, 35°
REFERENCE: INJ-S3-20/30/35, V-INJ-S2-30/35
EXPIERY DATE: 15.09.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.10.2017

