

CERTIFICATE OF ANALYSIS

Lot Number: 710HO30
PRODUCT V-ICSIPET HOLDING 30°
Product Number V-HOLD-30
Expiry Date: 14.10.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.11.2017

